

AIRLITEC Sarl 88, rue Jean Jaures 80470 Dreuil Les Amiens France  
 TEL.: +33 (0) 3 22 54 83 47 FAX: +33 (0) 3 22 54 83 29  
 Siret: 514 488 105 00016 Code APE: 4669B N° TVA: FR36514488105

To : SCHMIDT Technology GmbH  
 Feldbergstrasse 1  
 D-78112 St. Georgen/Germany

DATE: 14-févr-12

Att.: Elena Repp  
Tel.: Telefon +49 (0) 77 24 / 89 90  
Fax: Fax +49 (0) 77 24 / 89 91 01  
Email: [e.repp@schmidttechnology.de](mailto:e.repp@schmidttechnology.de)  
Web: <http://www.schmidttechnology.de>

Customer reference No. : 1202306  
 Our Order No. : 1202RH014  
 Contact person : Mr. Regis Houllier  
 Tel: +33 3 22 54 83 47  
 Fax: +33 3 22 54 83 29  
 Email: [regis.houllier@airlitec.com](mailto:regis.houllier@airlitec.com)  
 Web: [www.airlitec.com](http://www.airlitec.com)

ITEM	MODEL	DESCRIPTION	Q'TY	List Price	Discount	U/PRICE (EURO)	AMOUNT (EURO)	LEAD TIME (Weeks)
1		Repair Flow sensor SS20.650 : 524500-21411201 Repair OFFER 10004405 Serial number of the sensor: SN 900000002 Sensor change and recalibration	1	720,00	-10%	648,00	648,00	1

Total	EURO	648,00
Minimum Charge	EURO	0,00
* Packing & Handling charges	EURO	0,00
Freight Charge	EURO	0,00
Sub-total	EURO	648,00
VAT 19,6%	EURO	
<b>Total</b>	<b>EURO</b>	<b>648,00</b>

**Shipping instruction:** DOS SANTOS F.I.M. ELEC.  
 14b, Rue Maurice Halbwachs  
 51100 REIMS  
 France  
 Attention: Mr Christian Brun  
 Reference to indicate on shipping note :1202306

**TERMS and CONDITIONS:**

**Trade Terms:** Franco  
**Payment Terms:** 30 days net  
**Partial Shipment:** Not allowed.

Regis Houllier  
 Director

