

AIRLITEC Sarl 88, rue Jean Jaures 80470 Dreuil Les Amiens France
 TEL.: +33 (0) 3 22 54 83 47 FAX: +33 (0) 3 22 54 83 29
 Siret: 514 488 105 00016 Code APE: 4669B

To : M. Övül AKYOL
 Electrical&Electronic Eng.

DATE: 04-mars-11

HTG / ENTEK TEKNİK A.Ş.
Att.: Cevizli Mah. Tansel Cad.
Tel.: No:18 Maltepe / İSTANBUL
Fax: tel: 0216-4598660
Email: fax: 0216-4598370
Web: www.entekteknik.com

Your reference No. :
 Our Quotation No. : A2011RH030
 Contact person : Mr. Regis Houllier
 Tel: +33 3 22 54 83 47
 Fax: +33 3 22 54 83 29
 Email: regis.houllier@airlitec.com
 Web: www.airlitec.com

ITEM	MODEL	DESCRIPTION	Q'TY	U/PRICE (EURO)	AMOUNT (EURO)	LEAD TIME (Weeks)
1	518 210-21153S	SS20.400 Flow range see details next page Pressure: 700 to 1300 Hpas Sensor speed: 0 to 1m/s Sensor length: 130mm	8	683,00	5 464,00	4
	Option: 527 320	LED display Flow conversion Sensor power supply 2 alarm relays	8	385,00		4
	505 911-2	Connection cable (5m) with female connector	8	65,00		4
	301 082	Press fitting, stainless steel, clamp ring PTFE	8	61,00		
	505 960	Programming Kit	1	510,00		4
Total				EURO	5 464,00	
Minimum Charge				EURO	0,00	
* Packing & Handling charges				EURO	0,00	
Freight Charge				EURO		
Sub-total				EURO	5 464,00	
VAT 19,6%				EURO		
Total				EURO	5 464,00	

REMARKS:

* Lead time may be changed depending on the condition of the outstanding orders at our factory side.

TERMS and CONDITIONS:

Trade Terms: Ex work France
Payment Terms: 30 days net
Minimum Order Amount: Euro 150 per order (Packing & Handling and Freight charges excluded).
Partial Shipment: Not allowed.
Validity: 60 days from quotation date.
Cancellation: Not allowed after your order is acknowledged.

(The Trade Terms are in accordance with Incoterms 2000.)

Regis Houllier
 Director

No	Gas	Flow Capacity (m3/h)	Outside Diameter (mm)	Wall Thickness (mm)	Pressure (Bar)	Temperature (°C)	Inside Diameter (mm)	Speed (m/s)	PF	Gas Factor	Sensor	Flow Speed Limit AIR (m/s)	Max. Flow Rate (m3/h) AIR	Max Flow (m3/h) 20mA
1	Natural gas	0,05	22,00	1,00	atm	10 - 30	20,0	0,07	0,796	0,774	518 210-11143	1	515	0,7
2	Butane	0,05	22,00	1,00	atm	10 - 30	20,0	0,10	0,796	0,543	518 210-11143	1	515	0,5
3	Propane	0,05	22,00	1,00	atm	10 - 30	20,0	0,08	0,796	0,670	518 210-11143	1	1380	0,6

$$\text{Flow (m3/h)} = \text{Speed (m/s)} * \text{PF} * \text{PI} * \text{R(mm)}^2 * 0.0036 * \text{GF}$$